




All dimensions are in mm

Documents

Application note

AN001 "Calibration Services"

Material and plating

Tooling parts

Spanner

Material

Stainless steel

Mechanical data

Torque setting

2.00 Nm $\pm$ 0.1 Nm (force applied in the middle of handle)

Width across flats

11 mm

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG
RF_35/09.14/6.2

Technical Data Sheet				Rosenberger				
F	Torque Wrench			74W021-000				
Declaration of calibration options Factory Calibration Standard delivery for this torque wrench includes a Factory Calibration. The Calibration Certificate issued reports individual mechanical calibration results, traceable to national / international standards. Accredited Calibration Not available. For further, more detailed information see application note AN001 on the Rosenberger homepage. Calibration interval Recommendation12 months Packing Standard1 pce in box Weight121.5 g/pce								
While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.								
Draft		Date	Approved	Date	Rev.	Engineering change number	Name	Date
Martin Moder		18.05.15	Florian Reiner	21.02.25	d00	25-0071	M. Gehl	21.02.25
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de					Tel. : +49 8684 18-0 Email : info@rosenberger.de			Page 2 / 2